

# CE Declaration of Conformity

<b>Manufacturer/ Supplier Information:</b>	<b>BioFire Defense, LLC</b> 79 West 4500 South, Suite 14 Salt Lake City, Utah 84107, USA Phone: 1-801-262-3592 <a href="http://biofiredefense.com/">http://biofiredefense.com/</a>
<b>Date of Issue:</b>	03 JUNE 2009

## BioFire Defense declares that the products:

<b>Product Name:</b>	RAZOR EX
<b>Description:</b>	Bio-detection instrument using real time PCR.

To which this declaration relates, meets the protection requirements laid down in Council Directive 73/23/EEC on low voltage equipment safety and 89/336/EEC on the alignment of the legal provisions of the Member States on electromagnetic compatibility (EMC).

To assess the product with regard to electromagnetic compatibility, the following relevant harmonized standards were applied:

<b>EN 55022: 2006</b> Information technology equipment – Radio disturbance characteristics – Limits and methods of measurement
<b>EN 55024: 1998</b> Information technology equipment – Immunity characteristics – Limits and methods of measurement
<b>EN 61000-3-2: 2005</b> Electromagnetic compatibility (EMC) – Limits – Limits for harmonic current emissions (Equipment input current ≤ 16 A per phase)
<b>EN 61000-3-3: 2002</b> Electromagnetic compatibility (EMC) – Limits – Section 3: Limitation of voltage fluctuations and flicker in low-voltage supply systems for equipment with rated current ≤ 16 A
<b>EN 61010-1: 2001</b> Safety requirements for electrical equipment for measurement, control and laboratory use - Part 1: General requirements

We hereby declare that the equipment named above has been designed to comply with the relevant sections of the above referenced specifications. The unit complies with all essential requirements of the Directives.



**Amber Bawden**  
*Director, Quality Assurance*



**David Nielsen**  
*Vice President, Product Development*